



Ref. Certif. No.

DK-181995-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching Power Supply

Name and address of the applicant

WUXI COSEL ELECTRONICS CO LTD
NO.82-9, DICUI ROAD, BINHU DISTRICT Wuxi, Jiangsu, 214072
China

Name and address of the manufacturer

WUXI COSEL ELECTRONICS CO LTD
NO.82-9, DICUI ROAD, BINHU DISTRICT Wuxi, Jiangsu, 214072
China

Name and address of the factory

WUXI COSEL ELECTRONICS CO LTD
NO.82-9, DICUI ROAD, BINHU DISTRICT Wuxi, Jiangsu, 214072
China

Note: When more than one factory, please report on page 2

☐ Additional Information on page 2

Ratings and principal characteristics

1) Input: 200-480Vac, 0.75A, 50-60Hz Output: 24Vdc, 3.2A
2) Input: 200-480Vac, 0.5A, 50-60Hz Output: 24Vdc, 2.1A

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

1)FDA75F-24-x
2)FDA50F-24-x

☒ Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

Evaluated to National or Group Differences for EU Group Differences, CA, JP, CH, US

☒ Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 62368-1:2023

As shown in the Test Report Ref. No. which forms part of this Certificate

E495829-A6037-CB-1 issued on 2026-06-17

This CB Test Certificate is issued by the National Certification Body



☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2026-06-17

Signature:

Thomas Wilson



Ref. Certif. No.

DK-181995-UL

Additional Model Detail(s):

FDA75F-24-x, FDA50F-24-x, (May be provided with suffix"-x". Where "x" may be any combination of "N1", "N3"only for FDA75F, "C", "L")

Evaluated to National Differences declared for:

EN IEC 62368-1:2024, EN IEC 62368-1:2024/A11:2024

Additional information (if necessary)



- ☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☒ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2026-06-17

Signature:

Thomas Wilson